



Test Assay Advisory: Thyroglobulin Antibody Sensitivity

Offering our clients state-of-the-art testing is part of CPL's ongoing commitment to excellence.

Dear CPL Client,

We have just received a notice from Beckman-Coulter, the manufacturer of our Thyroglobulin Antibody assay reagent, that a possibility exists for decreased sensitivity for Thyroglobulin Antibody detection in a subset of patient samples tested with specific lot numbers of reagent. CPL was provided with the impacted lot numbers to test patients during the time period of 11/4/2018 to 2/19/2019. CPL immediately replaced this reagent lot with a new formulation not affected by this issue.

With potentially decreased sensitivity for TG Antibody, the impacted results may be inconsistent with clinical presentation, other laboratory tests, or imaging results. If you suspect that this issue impacted your patients' results, please contact your CPL Account Representative to review and arrange for reassessment. For clinical questions regarding this testing, please contact customer service at 800-633-4757.

We regret the inconvenience that this may cause you and your patients.

Thank you for supporting Clinical Pathology Laboratories